

**AGAGE\_scale\_2018\_v1 (updated in Jun. 2018)****Standard scales used in archived species from the measurement of AGAGE GC-MD and GC-MS instruments**

species	formula	scale	units	comments
methane	CH <sub>4</sub>	Tohoku University	ppb	
nitrous oxide	N <sub>2</sub> O	SIO-16	ppb	new SIO-16 scale used in 6/2017
carbon monoxide	CO	CSIRO94	ppb	
hydrogen	H <sub>2</sub>	MPI-2009	ppb	

**CFCs**

CFC-11	CCl <sub>3</sub> F	SIO-05	ppt	
CFC-12	CCl <sub>2</sub> F <sub>2</sub>	SIO-05	ppt	
CFC-13	CCIF <sub>3</sub>	METAS-2017	ppt	new compound released in 10/2017
CFC-113	CCl <sub>2</sub> FCCIF <sub>2</sub>	SIO-05	ppt	also includes the minor isomer CCl <sub>3</sub> CF <sub>3</sub> (CFC-113a)
CFC-114	CCIF <sub>2</sub> CCIF <sub>2</sub>	SIO-05	ppt	also includes the minor isomer CCl <sub>2</sub> FCF <sub>3</sub> (CFC-114a)
CFC-115	CCIF <sub>2</sub> CF <sub>3</sub>	SIO-05	ppt	

**HCFCs**

HCFC-22	CHClF <sub>2</sub>	SIO-05	ppt	
HCFC-141b	CH <sub>3</sub> CCl <sub>2</sub> F	SIO-05	ppt	
HCFC-142b	CH <sub>3</sub> CCIF <sub>2</sub>	SIO-05	ppt	

**HFCs**

HFC-23	CHF <sub>3</sub>	SIO-07	ppt	
HFC-32	CH <sub>2</sub> F <sub>2</sub>	SIO-07	ppt	
HFC-125	CHF <sub>2</sub> CF <sub>3</sub>	SIO-14	ppt	
HFC-134a	CH <sub>2</sub> FCF <sub>3</sub>	SIO-05	ppt	
HFC-143a	CH <sub>3</sub> CF <sub>3</sub>	SIO-07	ppt	
HFC-152a	CH <sub>3</sub> CHF <sub>2</sub>	SIO-05	ppt	
HFC-227ea	CF <sub>3</sub> CHFCF <sub>3</sub>	SIO-14	ppt	
HFC-236fa	CF <sub>3</sub> CH <sub>2</sub> CF <sub>3</sub>	SIO-14	ppt	
HFC-245fa	CHF <sub>2</sub> CH <sub>2</sub> CF <sub>3</sub>	SIO-14	ppt	
HFC-365mfc	CH <sub>3</sub> CF <sub>2</sub> CH <sub>2</sub> CF <sub>3</sub>	SIO-14	ppt	
HFC-43-10mee	CF <sub>3</sub> (CHF) <sub>2</sub> CF <sub>2</sub> CF <sub>3</sub>	SIO-14	ppt	

**Halons**

H-1211	CBrClF <sub>2</sub>	SIO-05	ppt	
H-1301	CBrF <sub>3</sub>	SIO-05	ppt	
H-2402	C <sub>2</sub> Br <sub>2</sub> F <sub>4</sub>	SIO-14	ppt	

**Chlorocarbons**

methyl chloride	CH <sub>3</sub> Cl	SIO-05	ppt	
dichloromethane	CH <sub>2</sub> Cl <sub>2</sub>	SIO-14	ppt	
chloroform	CHCl <sub>3</sub>	SIO-98	ppt	
methyl chloroform	CH <sub>3</sub> CCl <sub>3</sub>	SIO-05	ppt	
trichloroethylene	CHClCCl <sub>2</sub>	UB-98	ppt	
perchloroethylene	CCl <sub>2</sub> CCl <sub>2</sub>	NOAA-2003B	ppt	

carbon tetrachloride	CCl <sub>4</sub>	SIO-05	ppt
<b>Bromocarbons</b>			
methyl bromide	CH <sub>3</sub> Br	SIO-05	ppt
<b>PFCs</b>			
PFC-14	CF <sub>4</sub>	SIO-05	ppt
PFC-116	CF <sub>3</sub> CF <sub>3</sub>	SIO-07	ppt
PFC-218	CF <sub>3</sub> CF <sub>2</sub> CF <sub>3</sub>	SIO-07	ppt
<b>Other fluorinated compounds</b>			
sulfur hexafluoride	SF <sub>6</sub>	SIO-05	ppt
sulfuryl fluoride	SO <sub>2</sub> F <sub>2</sub>	SIO-07	ppt
nitrogen trifluoride	NF <sub>3</sub>	SIO-12	ppt

Notes:

### 1. The SIO-16 N<sub>2</sub>O Calibration Scale (June, 2017)

The SIO-16 calibration scale for N<sub>2</sub>O is based on a suite of 17 primary standard mixtures: 6 covering the 297-322 ppb concentration range that were also the basis of the SIO-98 calibration scale, and 11 covering the 310-354 ppb concentration range that were prepared for this new scale. Each primary standard was prepared by diluting high-precision pure N<sub>2</sub>O/CO<sub>2</sub> mixtures prepared manometrically in the Keeling CO<sub>2</sub> laboratory at SIO. CO<sub>2</sub> concentrations were measured in the resulting primary mixtures, referenced to Keeling laboratory CO<sub>2</sub> standards, by GC-FID with catalytic conversion to CH<sub>4</sub> (Weiss, J. Chrom. Sci., 19, 611-616, 1981) to determine N<sub>2</sub>O prepared values from the prepared N<sub>2</sub>O/CO<sub>2</sub> ratios. The optimal transfer from the Keeling CO<sub>2</sub> calibration scale was established from 9 reference cylinders based on Keeling manometric and optical measurements and improved CG-FID measurements with improved nonlinearity fitting. Uncertainties in this CO<sub>2</sub> scale propagation are at the < 0.1 ppm CO<sub>2</sub> (< 0.025%) level, and are subject to future revisions based on ongoing additional manometric measurements in the Keeling laboratory.

N<sub>2</sub>O concentrations in these 17 primary standards were measured against each other by GC-ECD (Prinn et al., J. Geophys. Res., 105, 17,751-17,792, 2000) and were fitted to a smooth curve of sensitivity vs. concentration to assign a "best estimate" N<sub>2</sub>O concentration (dry air mole fraction) to each standard mixture. The relative standard deviation of the corrections applied to the 17 individual prepared values is 0.017%.

The resulting SIO-16 N<sub>2</sub>O primary calibration scale was then propagated through the AGAGE "R1 scale" consisting of tanks of compressed whole air (Miller et al., Anal. Chem., 80, 1536-1545, 2008), to the entire AGAGE N<sub>2</sub>O atmospheric record from the beginning of the use of the R1 scale in AGAGE. Changes were also made in how the results were calculated: 1) The nonlinearity correction was changed to take concentration into account, instead of sample/standard ratio, and; 2) The new N<sub>2</sub>O scale was propagated to the R1 scale using a revised GC-ECD nonlinearity which has been constant since 2004, rather than the nonlinearity determined in 1998 that was used in the earlier propagation.

The resulting new AGAGE global atmospheric N<sub>2</sub>O values reported on the SIO-16 calibration scale have risen gradually compared to those reported previously, by from 0.0 ppb to approximately +0.8ppb over 20 years (+0.04 ppb/year). Approximately 20% of this increase is due to the use of the concentration based nonlinearity propagation, and about 80% of this increase is due to the use of the post-2004 GC-ECD non-linearity measurements. Importantly, there was no evidence of drift in the 6 original SIO-98 primary standards, and had the new calculation methods described above been used to propagate the SIO-98 scale, the corrections to present-day values, even though they are ~8 ppb above the range of these older standards, would have been much smaller than the changes reported here.

### 2. Comments for CFC-113 and CFC-114

Measurements of CFC-113 are a combination of CFC-113 (1,1,2-Trichlorotrifluoroethane, CAS 76-13-1) and CFC-113a (1,1,1-Trichlorotrifluoroethane, CAS 354-58-5). Measurements of CFC-114 are a combination of CFC-114 (1,2-Dichlorotetrafluoroethane, CAS 76-14-2) and CFC-114a (1,1-Dichlorotetrafluoroethane, CAS 374-07-2), see Vollmer et al., ACP, 2018